


Semi-continuous Benzene Monitoring in Burlington, Vermont



10/18/2007



Robert C. Lacaillade
Vermont APCD Air Toxics Project Manager

Overview:

Syntec Spectras GC955 Series 600 BTEX Analyzer

- General Information
- Design Specifications
- Principles of Operation
- Installation
- Operation
- Calibration
- Auditing
- Data handling/validation
- Quality assurance
- Results
- Issues/Concerns: "Bumps in the Road"




General Information

Syntec Spectras GC955 Series 600

BTEX Analyzer

- Manufacturer: Synspec
 - Located in the Netherlands
- US Distributor/Customer Support
 - Wilbur Technical Services, LLC (NH)
- Base unit cost ~ \$35k
- Burlington first ambient installation in USA
 - >800 Analyzers deployed in many EU countries to meet ambient Benzene monitoring Requirements
- Computer controlled-embedded PC running proprietary software
 - *Windows NT*, 40 Gigabyte HD
 - 10" integrated monitor
 - External keyboard & mouse
 - Various communication/data options (USB, modem, ethernet etc..)
 - *PC-Anywhere* (remote control)
- Standard 19" rack mount/110v




Design Specifications



GC955 Inside Top view


- Specifically for ambient air measurement of benzene, toluene & xylenes (others cpds/configurations available)
- Carrier gas- N₂
 - Ultra High Purity (Grade 5, 99.999%)
 - 50 psig, 1.5 ccpm
- Preconcentrator-Tenax GR
- 2-phase GC capillary column (in oven)
 - AT-5, .32mm ID
 - 2-meter "stripper" column isolates BTEX compounds
 - 13-meter "analysis" column
- PID- 10.6 eV, 50 µl measurement cell
- "Semi-continuous"; 15-minute run time
- Analog or digital data options
- Range: up to 300 ppbv
- Synspec's reported detection limit
 - benzene: 0.03 ppbv (0.1 µg/m³)

Principles of Operation



- 2 main operating modes:
 - Sample Injection
 - Sampling/Analysis
- Design keys
 - 10-port valve, preconcentrator and sampling piston/cylinder
 - allows for simultaneous collection of next air sample while analysis of previous air sample is occurring


Sample Injection Mode



- Sample collected on preconcentrator during previous 15-minute period is thermally desorbed into stripper column and then into analysis column
- Bypass pump operates for brief period before next sample collection
 - Flush sample piston/cylinder and SS sample inlet line
 - provide fresh air sample to "T" at back of analyzer

Sampling/Analysis Mode


- During each 15-minute run....
 - Sampling cylinder/piston draws total ~35 cc of ambient air sample through preconcentrator at set intervals
 - ~150 cc total during 15-minute cycle
- Once BTEX cpds make it thru stripper column (after injection mode).....
 - 10-port valve switches and back flushes remaining VOCs from stripper column.....
- While BTEX compounds continue on thru analysis column to the PID for measurement at specific retention times



Installation

- Rack mounted in monitoring trailer
- Connected to common glass manifold
 - 1/8" SS tubing
 - 2 μ sintered SS inlet filter
- 3 Channel analog output connected to ESC data logger
- Internet connection via dedicated DSL line
 - Allows offsite communication, data review/download
 - Remote control via PC-anywhere
- LAN connection with trailer PC

Operation




- Analyzer is pre-configured at factory for application/method
 - Run time, retention times, GC oven program, carrier gas flow, peak integration style etc....
- Easy to operate:
 - GC955 auto-starts run on power-up
 - Windows/menu-based operation
- Main run window indicates chromatographic results for current run
 - Retention times, conc., peak area
- Also displays real-time operating parameters at bottom
 - N₂ Flow & pressure, preconcentrator & oven temp, PID, sample pressure
- Pull-down menus provides access to:
 - Calibration results
 - Operational state
 - Configuration settings

Calibration

- Desired Range: 0-5 ppbv
 - Software has integrated calibration "mode"
 - results are stored and used to calculate current calibration relationship for each VOC
 - Linear and non-linear regression options available
 - Software has optional "auto-linearization" function for addressing non-linear PID response at lower levels.
- Retention times verified/edited
- BTEX cylinder gas: 1ppmv – *Spectra Gas*
- Onsite Zero Air System: *TEI Model 111*
- Diluter: *EnviroNics Model 6103* (trace level)
 - 0-10 sccm gas MFC; 0-20 lpm Air MFC



Benzene Calibration




- 6-point calibration
 - ppbv input levels 0.1, 0.25, 0.5, 1, 2, 4
 - Multiple 15-minute runs at each input level incorporated into curve
- Non-linear option used as best fit for desired range and accuracy at low level (auto-linearization for PID not activated at this time)
 - $R^2 = 0.9993$
 - y-intercept becomes lowest reportable level (0.13 ppbv \approx lowest calibrant)

Data Handling


- GC run time is synchronized with 15-min computer clock intervals (optional)
- Chromatogram results from each 15-minute sample run saved to unique data file
 - Saved files can be loaded and reviewed in main screen for review and final validation
- ppbv results for each 15-minute sample are written to a text file
 - Each text file contains 1 month of data
 - also contains RT and peak area
 - Text file is retrieved using PC-Anywhere
 - Other options are flash drive or FTP from trailer LAN
- Text file imported to *EXCEL* and *ACCESS* for review, processing, validation etc.....
 - 15-min values used to generate valid 1-hour averages
 - 75% data averaging rule used (minimum of 3, 15-min. values)
- Analog output to ESC datalogger provides external backup DAS option (limited to 3 compounds)
 - Provides real time access via network and additional flagging capability

Data Validation

- Daily review of hourly averages in ESC
 - Verify flags for maint./auditing
 - Assess data representativeness
 - Investigate missing, suspect or elevated values
- Review audit field sheets for performance w/in DQOs
- Text file of 15-min results reviewed and download to Excel/Access for:
 - processing, further review, flagging, generation of 1-hour avgs., validation
- Review chromatograms of randomly selected runs and ones for suspect or elevated values
 - Verify retention time, peak area, concentration
- Review results graphically
 - Time series, Fingerprint plots
 - Scatter Plot of benzene/toluene ratios for review of suspect values:



**Synspec GC955 Benzene vs. Toluene
Burlington 1-hour Avgs. from 6/20/07-8/31/07
(ppbv, n=1544)**



Quality Assurance

- QAPP approved by EPA-New England
- Weekly precision audit
 - Mid-range input level; DQO +/-15%
- Monthly zero/span check
 - 80-100% of range input level; warning: +/-10%, control: +/-20%
 - Zero: \leq lowest reporting limit (y-intercept)
- Biannual accuracy audit (second source standard)
 - Multiple mid-range inputs; DQO +/-20%
- Data capture
 - Quarterly DQO: 75%

Quality Assurance Results

(6/20/07- 8/31/07)


Currently, DQO's Are Being Met

Mean precision error:

- Benzene=-6.5%
- Toluene=-2.5%

Data Capture

- 15-minute \approx 87%




Benzene & Toluene 1-hour Averages*

6/20/07-8/31/07




**Burlington Benzene and Toluene 1-hour Avgs.
from 6/20/07- 8/31/07 (ug/M3)**



* final review/validation pending; 1-hour averages only flagged for missing and audit/maintenance periods

One week in August with confirmed gasoline delivery near site


Burlington Benzene and Toluene 1-hour Avgs.
from 8/26/07- 8/31/07 (ug/M3)





■ Similar period in September with confirmed gasoline delivery near site

← Visual confirmation of tanker delivery from onsite video cam



9/14/07 3:23

← Pollution rose confirming S/SSE winds during highest values




← Benzene time series (ESC) showing elevated levels for Hr 0300-0400

Burlington BTEX Results vs. TO-15 canisters

(Comparison of 24-hour Averages for Benzene and Toluene)

**BTEX 24-hr averages based on mean of 96 separate 15-minute values for each day*
**TO-15 results based on analysis of 24-hour integrated whole-air canister sample*

Comparison of GC955 BTEX Results to TO-15 Canisters Benzene & Toluene 24-hour Averages, Burlington VT (ppbv)



- ◇— Benzene 6-liter 24-hr sample
- Benzene 6-liter 24-hr sample dup
- BTEX Benzene (24-hour avg. of 15-min values)
- ◇— Toluene 6-liter 24-hr sample
- △— Toluene 6-liter 24-hr sample dup
- BTEX Toluene (24-hour avg. of 15-min values)

"Bumps in the Road"

- Computer/hardware issues
 - Computer reboots periodically
 - GC955 Analyzer occasionally locks-up due to a Windows error
 - Can create large missing data periods
 - Still under investigation; computer replacement considered
 - Analog output only provides for 3 channels
 - Calibration/audit limitations
 - 0-10 ccpm MFC a necessity
 - Canisters not viable option
 - Tedlar bags ok, not preferred option
- BTEX software issues
 - European date format in data text file requires additional post processing
 - Chromatogram requires *Synspec* software to review and is not a permanent record of results for ppbv value
- Communication/Customer Support
 - Some difficulty/delays dealing with the Netherlands regarding timely corrective actions and answers

Thank you!

- Jenny Berschling (Vermont APCD/DEC Lab)
- John Simone (Vermont APCD)

- John Wilbur (Wilbur Technical Services)
 - For all their effort and invaluable assistance with this project

Contact Information

- Note: Synspec will be in the U.S. next month:
 - What: **North American Technical Training Session**
 - When: **November 7-9, 2007**
 - Where: **Sheraton in Braintree, Massachusetts**
 - Contact: **Wilbur Technical Services**
- Wilbur Technical Services, LLC
 - John Wilbur (603) 880-7100
- Synspec
 - Michael Rijpkema, Wouter Lautenbach
 - Website: www.synspec.nl
 - Phone: (011) 31 50 526 6454 (+6 hours)
- Robert Lacaillade, Vermont APCD
 - Email: robert.lacaillade@state.vt.us
 - (802) 241-3840